

Calibration Certificate

Customer Address:

Schneider Electric
3700 6th St. SW
Cedar Rapids, IA 52404
319-369-4631

Calibration Laboratory Address:

Schneider Electric
3700 6th St. SW
Cedar Rapids, IA 52404
319-369-4631

REVIEWED

By Kevin Woods at 4:07 pm, May 07, 2020

Control # 063-190-4
Test Results Pass
Technician Dan Ginter
Temperature 22 °C
Humidity 37 RH
Data Type FOUND LEFT
MU File MU-063M
Instrument Condition Operational

Customer # 063-190-4
Location ENDURANCE FACILITY
Manufacturer HBM
Model Number 1-GN610B-2
Description DIGITAL OSCILLOSCOPE
Serial Number ILB1700943 CH 4
Cal Date 4/8/2020
Due Date 4/8/2021

Certificate Number: 2020000963

Procedure Name HBM GEN SERIES DATA RECORDER - 5720, 3458

Work Order Notes: CAL'D TO MFR SPEC. DATA LOCATED: CONTROL#'S/063-XXX/2020/063-190 DATA ALL CHANNELS.PDF

Schneider Electric certifies that the instrument described above meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). It has been calibrated using measurement standards traceable to the SI through the National Institute of Standards and Technology (NIST), or by using natural physical constants, intrinsic standards or ratio calibration techniques. This laboratory conforms to ISO 9001:2015 ISO/IEC 17025:2017 Measurement uncertainty will be available in a reasonable amount of time upon written request. Accessing measurement uncertainty files, see document ALL-001 located in the Quality Management System. Estimated expanded uncertainty is computed at 95% confidence level, coverage factor k=2 following the Met-002 procedure. Received date of the calibration items will be entered when it is critical to the validity of the results.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the organization issuing this report.

Standards Used

Control #	Manufacturer	Description	Model Number	Cal Date	Due Date
018-270	AGILENT	DMM, 8 1/2 DIGIT	3458A	8/14/2019	8/14/2020
083-016	FLUKE	CALIBRATOR	5730A	3/16/2020	3/16/2021

HBM GEN SERIES DATA RECORDED						
MET/CAL Results		Pass		FOUND_LEFT		WTUSLVSE114619D
Test Description	True Value	Lower Limit	Test Results	Upper Limit	Status	TSR
DOES THIS PROCEDURE CONFORM TO ISO/IEC 17025						
Result of Operator Evaluation					Pass	
TEMP. AND HUMIDITY WITHIN REQUIREMENTS?						
Result of Operator Evaluation					Pass	
UNIT PASSED CALIBRATION AND DATA PDF'D TO C/N DIRECTORY?						
Result of Operator Evaluation					Pass	
-- End of measurement results--						