



9-26-2008

Efficacy of an ankle orthosis with a subtalar locking system in restricting ankle kinetics and kinematics in lateral cutting

Songning Zhang

University of Tennessee - Knoxville, szhang@utk.edu

Michael Wortley

University of Tennessee - Knoxville

Qingjian Chen

University of Tennessee - Knoxville

Julia Freedman

University of Tennessee - Knoxville

Casey Riley

University of Tennessee - Knoxville

Follow this and additional works at: http://trace.tennessee.edu/utk_biopubs



Part of the [Medical Biotechnology Commons](#)

Recommended Citation

Journal of Foot and Ankle Research 2008, 1(Suppl 1):O19 doi:10.1186/1757-1146-1-S1-O19

This Presentation is brought to you for free and open access by the Division of Biology at Trace: Tennessee Research and Creative Exchange. It has been accepted for inclusion in Faculty Publications and Other Works -- Division of Biology by an authorized administrator of Trace: Tennessee Research and Creative Exchange. For more information, please contact trace@utk.edu.

Oral presentation

Open Access

Efficacy of an ankle orthosis with a subtalar locking system in restricting ankle kinetics and kinematics in lateral cutting

Songning Zhang*, Michael Wortley, Qingjian Chen, Julia Freedman and Casey Riley

Address: Biomechanics/Sports Medicine Lab, The University of Tennessee, Knoxville, TN, USA

Email: Songning Zhang* - szhang@utk.edu

* Corresponding author

from 1st Congress of the International Foot & Ankle Biomechanics (i-FAB) community
Bologna, Italy. 4–6 September 2008

Published: 26 September 2008

Journal of Foot and Ankle Research 2008, 1(Suppl 1):O19 doi:10.1186/1757-1146-1-S1-O19

This abstract is available from: <http://www.jfootankleres.com/content/1/S1/O19>

© 2008 Zhang et al; licensee BioMed Central Ltd.

Introduction

The ankle joint is the most injured joint during sports participation [1]. Ankle orthoses have been shown to be effective in reducing ankle inversion injuries and are often prescribed for rehabilitation and prevention of lateral ankle sprains. Efficacy of ankle orthoses is often assessed by comparing reduction of passive inversion ROM as well as ankle kinematics between braced and unbraced movements [2,3]. However, joint kinetic responses in lateral cutting were rarely examined. Therefore, the objective of this study was to examine the effectiveness of a new semi-rigid ankle orthosis with a subtalar joint locking mechanism in restricting ankle kinetics and kinematics during a lateral cutting movement.

Methods

Ten female and ten male subjects performed five lateral cutting trials in each of four conditions wearing no brace (NB), a semi-rigid Element ankle brace with a calcaneal

and subtalar locking system (AB1, DeRoyal), a semi-rigid Functional ankle brace with a hinge joint (AB2, DeRoyal), and a soft ASO lace-up ankle brace (AB3, Medical Specialties). A seven-camera motion analysis system (240 Hz, Vicon Motion Analysis Inc.) and a force platform (2400 Hz, AMTI) were used to obtain the three-dimensional kinematics and ground reaction force data respectively. A one-way repeated measures ANOVA was used to evaluate differences among the brace conditions on selected variables ($p < 0.05$) with post hoc comparisons conducted to detect specific differences among the braces using a Bonferroni adjustment (SPSS, Inc.).

Results

For the angular velocity, the peak contact inversion velocity (On_Y) was significantly reduced for AB1 compared to the control group (Table 1). No significant differences were seen in the peak lateral impact GRF (Min_X) among the brace conditions. However, the peak vertical GRF

Table 1: Selected kinematic and kinetic variables: Mean ± STD.

Cond	On_Y (deg/s)	Min_X (BW)	Max_Z (BW)	Min_Y (Nm/kg)
NB	322.3 ± 103.7 ¹	-1.05 ± 0.24	1.71 ± 0.29 ¹	-0.58 ± 0.19
AB1	263.6 ± 113.6	-0.98 ± 0.21	1.58 ± 0.20 ³	-0.60 ± 0.16
AB2	257.5 ± 128.6	-1.02 ± 0.27	1.67 ± 0.28	-0.62 ± 0.19
AB3	295.7 ± 129.9	-1.04 ± 0.23	1.68 ± 0.24	-0.60 ± 0.18

¹ significantly different from AB1, ³significantly different from AB3.

(Max_Z) for AB1 was significantly smaller than no brace and AB3. In addition, the peak ankle eversion joint moment (Min_Y) did not show significant differences among the brace conditions (Table 1).

Conclusion

The ankles did not reduce peak horizontal GRF data which is consistent with the findings of Cordova and his colleagues [2] in a shuffle movement. However, the peak vertical GRF was reduced with the Element brace compared to the ASO and control condition. Furthermore, the results showed that the Element ankle brace provides restriction of ankle inversion at early contact and pushoff. These results suggest that the Element brace is more effective in the lateral cutting. In addition, the tested orthoses also accommodate movement requirements that are commonly desired of an effective ankle orthosis.

Acknowledgements

Funded by DeRoyal Industries, Inc.

References

1. Hootman JM, et al.: *J Athl Train* 2007, **42(2)**:311-319.
2. Cordova ML, et al.: *Med Sci Sports Exerc* 1998, **30(9)**:1363-1370.
3. Gross MT, et al.: *J Orthop Sports Phys Ther* 1997, **25(4)**:245-252.

Publish with **BioMed Central** and every scientist can read your work free of charge

"BioMed Central will be the most significant development for disseminating the results of biomedical research in our lifetime."

Sir Paul Nurse, Cancer Research UK

Your research papers will be:

- available free of charge to the entire biomedical community
- peer reviewed and published immediately upon acceptance
- cited in PubMed and archived on PubMed Central
- yours — you keep the copyright

Submit your manuscript here:
http://www.biomedcentral.com/info/publishing_adv.asp

